# UAB GE MRI Scanners

| Scanner Name | PETMR | GARDMR | LEEDSMR | TKCMR4 | CVMR1 |
|---|---|---|---|---|---|
| Scanner Location | Advanced Imaging Facility | Gardendale Hospital | Leeds Hospital | The Kirklin Clinic | Boshell Building |
| Scanner Contact | Lael Gore | Gail Reato | Kayla Freeman | Heather Edmonds | William Fisher |
| Phone Number | (205) 996-3272 | 205-996-3565 | (205) 699-4180 | 205-801-7901 | (205) 934-9906 |
| System ID | UABMRPET | 205608MR | 205702450W | 205801KMR3 | 205934CVMR |
| Magnetic Field Strength | 3.0T | 1.5T | 1.5T | 1.5T | 1.5T |
| Scanner Model | PETMR | Artist | Optima 450w | Excite TwinSpeed | EchoSpeed Cvi |
| System Serial Number | UA0408 | HM1901 | HM1181 | R2537 | R0233 |
| Installation Date | 1.2015 | 8.2018 | 7.2015 | 3.2000 | 3.1999 |
| Software Revision | MP26.0 | DV26.0 | DV26.0 | HD16.0 | LX9 |
| Gradient Type | XRMw | XRMw | XRMw | TwinSpeed | EchoSpeed |
| Gradient Strength (mT/m) | 44 | 44 | 44 | 33 | 33 |
| Slew Rate (mT/m/msec) | 200 | 200 | 200 | 80 Whole / 120 Zoom | 120 |
| Number of Receivers | 32 | 96 | 32 | 16 | 4 |
| Multi-Nuclear (MNS) Installed | N/A | N/A | N/A | N/A | N/A |
| Bore Diameter (cm) | 60 | 70 | 70 | 60 | 60 |
| Accreditation Program (MRAP) | 54432 - 01 | 54520 - 01 | 53666 - 01 | 03698 - 04 | IACMR |
| MRAP Expiration Date | 3/8/2022 | 8/22/2022 | 5/2/2023 | 9/28/2024 | 3/31/2022 |
| Breast MRAP (BMRAP) | N/A | 51274 - 01 | 51340 - 01 | 00677 - 01 | N/A |
| BMRAP Expiration Date | N/A | 6/20/2022 | 1/16/2023 | N/A | N/A |