

UAB Philips MRI Scanners

Scanner Name	UABMR1	UABMR2	TKCMR1	TKCMR2	TKCMR5	CVMR2
Scanner Location	North Pavilion	North Pavilion	The Kirklin Clinic	The Kirklin Clinic	The Kirklin Clinic	Boshell Building
Scanner Contact	Chris Horsley	Chris Horsley	Heather Edmonds	Heather Edmonds	Heather Edmonds	William Fisher
Phone Number	(205) 975-3771	(205) 975-3771	205-801-7901	205-801-7901	205-801-7901	(205) 934-9906
System ID	53621136	53621167	535208	535211	61297506	104144
Magnetic Field Strength	1.5T	3.0T	1.5T	3.0T	1.5T	3.0T
Scanner Model	Ingenia	Ingenia	Achieva	Achieva	Ingenia	Achieva
System Serial Number	41034	42033	20126	17108	70051	17093
Installation Date	7.2010	7.2010	3.2003	3.2003	3.2017	1.2000
Software Revision	R5.3	R5.3	R3.2	R3.2	R5.4	R3.2
Gradient Type	Omega HP	Omega HP	Quasar Dual	Quasar Dual	Omega HP	Quasar Dual
Gradient Strength (mT/m)	45	45	33/66	33/66	45	66
Slew Rate (mT/m/msec)	200	200	90 Whole / 180 Zoom	90 Whole / 180 Zoom	200	90
Number of Receivers	unlimited	unlimited	16	16	unlimited	16
Multi-Nuclear (MNS) Installed	N/A	N/A	N/A	N/A	N/A	N/A
Bore Diameter (cm)	70 cm	70 cm	60 cm	60 cm	70 cm	60 cm
Accreditation Program (MRAP)	N/A	N/A	03698 – 06	03698 – 05	03698 – 07	IACMR
MRAP Expiration Date	N/A	N/A	9/28/2024	9/28/2024	9/28/2024	3/31/2022
Breast MRAP (BMRAP)	N/A	N/A	N/A	N/A	00677 - 02	N/A
BMRAP Expiration Date	N/A	N/A	N/A	N/A	N/A	N/A