

## **MRI MSK Exam Post Orthopedic Hardware Procedure**

Technologist are allowed to safely image patients immediately after surgery containing orthopedic hardware, including plates, pins, staples, rods, plates, and screws at all field strengths.

If the patient has an orthopedic implant, the technologist must check, verify the orthopedic hardware, and document for MRI compatibility. The technologist is to consult the Radiologist for approval if documentation is not available.

If you have any questions or concerns, please contact the Radiologist.



**Sam Galgano, MD | Associate Professor, Vice Chair of Quality and Patient Experience**

Section Chief, Abdominal  
Imaging University of Alabama at  
Birmingham Department of  
Radiology